This is a mirror page of http://www.seeingwithsound.com/neuralink.htm

White paper: Why The vOICe will defeat Neuralink Blindsight

PDF version

« The vOICe Home Page

Why The vOICe vision BCI for the blind can and will defeat Neuralink Blindsight and other brain implants for restoring vision

Neuralink Blindsight and most other brain implants aimed at restoring vision to the blind work by implanting electrodes in (or on) the primary visual cortex. A small current through an electrode can give a visual percept in the form of a phosphene, a flash of light. The assumption is that by stimulating a grid of many electrodes one can evoke the perception of an image, and thereby restore a form of vision. The vOICe visual-to-auditory sensory substitution approach instead uses normal hearing to convey images encoded in sound, and the human brain must first learn to decode these sounds to obtain a form of vision through sound-guided mental imagery. In the following, we will focus mostly on comparing The vOICe with Neuralink Blindsight, but much of what is said about the latter also applies to other brain implants in the visual cortex. The vOICe approach too is not without limitations and unknowns, but overall has many key advantages. To scientists: we greatly welcome your feedback, insights and particularly studies!

	Neuralink Blindsight	The vOICe vision BCI
Invasive or non-invasive?	Invasive: requiring brain surgery with risks of infection, hemorrhage, seizures and death, as well as inevitably some damage to the cortical tissue where electrodes are inserted, both physical damage from the act of insertion and from possible subsequent inflammatory reactions and gliosis. Invasive BCIs can be characterized by a dimensionless "Butcher number" (Meister, 2019): the number ratio of damaged or killed neurons and stimulated or recorded neurons. Optogenetic approaches that avoid inserting electrodes introduce other risks, by for instance using viruses (viral vectors) to deliver genes into cells, potentially increasing cancer risk through mutations, while those genes themselves may trigger immune responses or prove cytotoxic through new protein expression. Safety limits for chronic use of current or charge injection with large sets of thin penetrating electrodes in human visual cortex are still unknown.	Non-invasive: no known medical risks beyond the avoidable risk of hearing damage when sounds are played at high volume for prolonged periods of time, like with loud music. No neurons are damaged or killed. Demonstration: reaching and grasping by Orion brain implant recipient (left) and by a blind user of The vOICe vision BCI fetching multiple objects (right).
Messy or orderly?	Messy: phosphenes (light flashes) at somewhat random-offset positions (like "stars-in-the-sky"), with variable sizes and shapes, sometimes flashing, with non-veridical colors and with gaps between sometimes partially overlapping phosphenes. Phosphene positions are further affected by representational drift, causing apparent shifts within the visual field that require frequent recalibration to undo. Quality of vision will be far worse than early Nintendo graphics or Atari graphics at first, and will in all likelihood never exceed or even approach normal human vision. Brain implants for vision might benefit from structure-from-motion information, but it is currently unclear to what extent human structure-from-motion perception depends on retinal motion processing that is unavailable with a brain implant.	Orderly: sounds appearing at perfectly orderly and uniform spatial and temporal positions through pitch (tone frequency), stereo position and time. Negligible (if any) representational drift. Example soundscape sequence for a virtual reality scene that shows a path lined with pillars, next approaching a wall with a gate, and with a sphere and cube behind the wall: VR 1 on YouTube. Use stereo headphones. Also VR 2 on YouTube, which uses a slightly different audio rendering. Even if you cannot properly visualize the soundscapes yet, you likely can confirm that all the key visual elements are audible "in hindsight". Next try to imagine what interpreting this animation would be like with the scattered phosphene blobs of a brain implant.
See light?	Yes: visual percepts (light flashes) right upon first-time use, giving a great honeymoon experience, but it does not necessarily yield interpretable images. Meaningful "vision" may or may not come much later after extensive practice and rehabilitation, and artificial vision results may disappoint given the hype. An analogy is what happened with the Argus II retinal implant of Second Sight Medical Products (out of production since 2019): initially, blind recipients were thrilled to perceive light and flashes of light again, but subsequently they never developed functional vision beyond mere light blob detection, but that finding was long after the initial mass media coverage. The only 2 blind Russian Argus II recipients were beaten by a blind user of The vOICe vision BCI glasses at the Neurothlon 2018 contest in Samara, Russia.	Not from the start: initially perceived as an annoying cacophony of "just sound". Visual percepts and vision may or may not arise from extensive practice including sound-guided mental imagery (cf. YouTube exercise video for the sighted, and use stereo headphones), while the sounds serving as visual information carrier are no longer or barely consciously perceived.  Later on these adaptations might be boosted and sped up through additional means such as non-invasive brain stimulation (rTMS, tDCS, tACS, tRNS, tUS/tFUS, ...) and drugs, depending on scientific progress in neuroscience. Some blind users have reported having visual percepts (cf. "Visual experiences in the blind induced by an auditory sensory substitution device").
Late-blind and congenitally blind?	Works only for late-blind people - no matter what Elon Musk claims. Congenitally (totally) blind people do not get visual percepts from electrically stimulating their "visual" cortex (occipital cortex), because neural plasticity caused it to be recruited for other, non-visual, purposes. There is zero evidence that visual percepts can still arise at a later stage in congenitally blind people. (An intermediate case is that of one late-blind Dobelle brain implant recipient, Dennis, who went blind first during adolescence, who initially heard beeps instead of seeing phosphenes upon stimulation of his visual cortex while his visual cortex got re-recruited for vision.)	Works for both congenitally blind and late-blind people. This assumes that they have more or less normal hearing, but some existing blind users are hard of hearing and use hearing aids, so this is a soft requirement. It does not require visual percepts to understand that a rising tone means a rising bright line, such that congenitally blind people can learn to use it as well to acquire at least some functional vision. For instance to locate and reach for a coffee mug on a table top without groping, to orient toward a doorway or reach for a shiny door handle, track a walking path, spot the stripes at a marked crosswalk (zebra crossing), etc., etc.
Textures and shading?	Cannot convey textures and shading, only a couple of lines and edges. Conveying textures, shading and uniform areas would require simultaneously activating many electrodes above phosphene threshold, but this would likely evoke seizures or could even cause death. E.g., activating 10,000 electrodes with a 10 µA (microampere) suprathreshold current per electrode would amount to a total current injection of 100 mA (milliampere). Therefore edge detection algorithms are first applied to the camera view to limit the maximum number of simultaneously active electrodes, making anything resembling normal vision impossible. When looking up at a clear blue sky you would probably see... nothing, because there are no edges.	Can convey textures, shading and smooth surfaces without any restriction. Shading and smooth surfaces give smooth noise-like sounds. Textures give characteristic complex dense "rhythms", and can be perceptually more useful than edges when following a path, where edges of surrounding objects or of sharp path-crossing shadows from buildings or trees can be very confusing.

Scalable?	Barely scalable. Expensive surgery with special implant hardware limits the potential market to a small number of rich countries, and even there further limited by reimbursement policies in health insurance, because few blind people have the financial resources to pay for the brain implant themselves. Surgery for a visual cortical prosthesis is further complicated by the fact that the visual cortex lies in part hidden within the banks of the calcarine sulcus. Elon Musk may in line with his general ambitions force the company to move on to greener (more lucrative, mass-market) pastures once the technology has been proven to work, somewhat.	Highly scalable at extremely low cost. The vOICe can already run on billions of existing devices: The vOICe for Windows on PCs, The vOICe for Android (APK file) on Android smartphones, tablets and smartglasses, and The vOICe progressive web app on "anything" with a modern browser, i.e., not only on PCs and Android devices, but also on iPhone and iPad (Safari browser), Linux devices, Raspberry Pi devices, etc. Many blind people, even in developing countries, already have at least a smartphone that is compatible with The vOICe. Smartphones can be made head-mounted using a cheap VR headset holding the phone. Not convenient, but highly affordable. Higher-end (more expensive) options for head-mounted use with compatible smart glasses already exist as well, such as the Vuzix Blade and Vuzix Shield smart glasses with The vOICe for Android app available from the Vuzix App Store, but the same app in principle runs on all stand-alone Android-based smart glasses with a camera, such as smart glasses from TCL RayNeo, DigiLens and ThirdEye.
Device lifetime?	Device lifetime is still unknown, likely only several years. Previous work with implanted penetrating electrodes in visual cortex showed effective lifetimes of up to several years (often less), but much depends on the specific proprietary electrode technology. Electrodes typically show signs of corrosion, cracks and other damage after prolonged use in vivo, with best results thus far reported using PEDOT electrode coatings. Penetrating electrodes may also retract and fail due to movement of the brain relative to the skull. Only the external components can be replaced without surgery, but may become obsolete once the company decides to no longer support the hardware. The implant itself cannot be replaced without yet another brain surgery, each time leading to some additional tissue damage in the visual cortex.	Device lifetime is effectively unlimited. Using globally available mature hardware such as smartphones and stereo earbuds, lifetime per hardware component is likely on the order of a decade or more, while replacing the hardware is easy and affordable. Some basic but fully functional versions of the software are open source as well as publicly archived in for instance the Internet Archive, so one may reasonably expect to be able to use the technology during one's whole (human) lifetime. This also means that independent scientific studies based on use of The vOICe are less likely to become outdated or irrelevant while companies and their products come and go. Musicians can have a long life performing at a high level if they take care about their hearing. Moreover, the platform-independent The vOICe web app will be around for the next 1,000 years (cf. 2020 badge) as part of a data repository preserved in the Arctic World Archive, located 250 meters deep in the permafrost of an Arctic mountain, stored in a steel-walled container inside a sealed chamber within a decommissioned coal mine. The vOICe web app is also in the Internet Archive.
Training support?	Unspecified support for rehabilitation training. "Vision" with a brain implant is totally different from normal vision or even low vision eyesight, so it will after recovery from brain surgery require extensive rehabilitation to learn to make good, practical and safe use of the brain implant in daily living situations.	The vOICe too requires extensive training for best results. There is still lack of 1-on-1 training support that would seem to best fit local blindness institutes already involved in training for safe blind use of the white cane and/or guide dog, and other daily living skills. An online English self-training manual for The vOICe is available, with translations into Russian, Portuguese and Chinese. Its use requires self-discipline, and results would greatly benefit from supplemental on-site feedback and continued motivation by sighted instructors, including existing O&M (orientation and mobility) instructors in your area. Suppose you'd need to train for even 5 years to adequately master The vOICe: you need to weigh that against the probability that there will be any brain implant for vision on the market within 5 years, its cost, its own training effort, and next the expected useful lifetime of that device, including continued company support.
Resolution?	Visual resolution is still unknown, but unlikely to benefit much from more than 1,000 - 10,000 electrodes in view of the fact that only lines and edges can be conveyed, to limit seizure risk. This would limit resolution to between 30 × 30 and 100 × 100 for a few lines and edges, but overlapping receptive fields may give a drastically lower effective resolution. Neuralink is reportedly planning to implant about 6,000 electrodes in their first BlindSight brain implant recipient.	 Visual-to-auditory conversion is by default done at a resolution of 176 × 64 pixels, equivalent to stimulating over 11,000 electrodes. The effective resolution may lie between about 30 × 30 and 60 × 60, in part depending on actual image content, but theoretically there are possibilities for higher resolution. However, even a resolution as low as 25 × 25 is already considered adequate for ambulatory vision. A visual acuity on the order of 20/160 to 20/240 seems achievable in combination with a field of view on the order of 60°, even without foveal enlargement. An acuity worse than 20/200 (or tunnel vision with a field of view smaller than 20°) defines legal blindness in the US, so an acuity slightly below or slightly above this acuity appears plausible. Importantly, unlike with human vision this acuity is uniform across the entire field of view, and in the visual periphery it can therefore even exceed normal vision.
Field of view?	The field of view (FOV) is still unknown, but likely mostly limited to a cortical area corresponding to the fovea, perceptually giving severe tunnel vision (1° to 2° visual angle, about the size of a thumbnail at arm's length). All image content is squeezed into this narrow field of view. Since the definition of legal blindness in the US applies when having a field of view smaller than 20°, the implant recipient would likely still count as legally blind no matter how high the visual acuity.	The field of view (FOV) is typically 60° to 120° depending mostly on the choice of camera. For all vision systems there is an inevitable trade-off between visual acuity and field-of-view, but as stated above a visual acuity on the order of 20/160 to 20/240 seems achievable in combination with a field of view on the order of 60°.
Frame rate?	Frame rate is still unknown, but neural recovery time after stimulation (refractory period plus perceptual fading upon repeated stimulation) may limit effective frame rate to, say, 4 to 8 frames per second (4-8 fps). Neural habituation (reduced neural response to repeated stimuli) may further substantially reduce effective frame rate and bandwidth. The need for extensive (head) scanning to compensate for a narrow field of view may defeat any frame rate advantage.	Frame rate is by default 1 frame per second (1 fps), optionally set to a higher 2 or 4 fps (or even higher) in a trade-off with effective visual resolution.
Masking?	Need not cause sensory masking, but this only applies to totally blind people. Visual cortex stimulation likely interferes with and degrades natural low vision, for instance because the field of view of both types of sensory input does not match, but other forms of detrimental interference seem likely.	Some degree of masking of hearing natural sounds is inevitable. Minimized by using a modest sound volume in relation to ambient sound levels, and by using bone conduction headphones that do not block echolocation and natural hearing. The gains in functional vision should outweigh auditory masking effects.

Ambiguity?	Unknown apart from general visual ambiguities and illusions, and the limitations of ultralow vision (ULV).	Very rare, and easily disambiguated. In rare cases, soundscapes of for instance horizontal line segments may sound similar to environmental sounds such as the beeping of a truck backing up. If you then look slightly up or down and the pitch changes, you know it is not a truck backing up.
Literature?	Virtually no peer-reviewed scientific literature. It is for believers or NDA-signers. Also note that the term "blindsight" as such has been in use for many years in the scientific and other literature to describe the phenomenon that people who are totally blind due to damage to their visual cortex, i.e. cortical blindness, sometimes still can "guess" better than chance about the presence or properties of visual items in front of them, thereby for instance avoiding obstacles through a kind of subconscious vision. This should make (always case-insensitive) trade-marking the term by Neuralink in the close context of a crude form of vision for the totally blind impossible. Similar comments apply to attempts to trademark for instance the terms telepathy and telekinesis.	Extensive peer-reviewed scientific literature. We welcome independent studies from all angles, without strings or NDAs attached, and yes, the learning curve and (for late-blind users) qualia barrier remain major concerns and uncertainties. To what extent is it possible to nudge even the adult human brain toward new modes of processing through functional (not physical) rewiring? Join the quest! To broaden the scientific scope one may embed the above in the many whole-brain dynamic and plastic functional connectivity questions that are now only beginning to be addressed. Perhaps grant-writing can even be fun... for instance think of mastering The vOICe as establishing a specific auditory-to-visual neural manifold through functional rewiring.
Patient/user reports?	Blind patient reports? Nothing yet at the time of writing.	Blind user reports in numerous places and contexts. For example, former Dobelle brain implant recipient Jens Naumann on YouTube (playlist), blind user Pat Fletcher on YouTube, congenitally blind user Pranav Lal on YouTube, Russian blind user Vadim Artsev on YouTube.
Deaf-blind users?	No known reason why a brain implant in visual cortex would not work for late-blind people who are also profoundly deaf.	The vOICe assumes more or less normal hearing, but is known to be also used by totally blind people who are hard of hearing, in combination with hearing aids. The vOICe will not work for people who are profoundly deaf, but may work to some extent for totally deaf-blind people who received a cochlear implant. There are no known reports about this yet.
AI support?	Unknown.	AI-based recognition of camera view content can be useful but is not always reliable, and is therefore considered supplemental and optional. The vOICe for Android includes support for offline real-time talking OCR (text recognition), face detection, basic object recognition and barcode & QR code recognition. More advanced third-party AI apps for object recognition and scene description (e.g. Microsoft Seeing AI or Google Lookout) can be launched from within The vOICe for Android.
Funding?	Essentially unlimited funding as long as Elon Musk believes in it. He is likely after bigger fish than the niche market of Neuralink Blindsight for totally blind people.	Available and sustainable indefinitely even without funding. There may come mass market applications in brain training and building cognitive reserve, but those are hypothetical for the time being and not required for continued availability. See also the section about device lifetime.
Animal experiments?	Experiments with non-human primates and other animals are required to establish safety and thereby gain regulatory approval for use in humans. In the past there has been a lot of controversy about Neuralink's animal experiments although Neuralink denied animal cruelty claims.	No animal experiments are required nor needed for safe deployment of The vOICe with humans. That said, it is possible that animal experiments are performed by independent institutes in the context of neuroscience studies on for instance crossmodal brain plasticity and functional rewiring.
Ethical issues?	The above and other aspects of brain implants raise numerous ethical concerns beyond the need for animal experiments. There can for instance be harmful medical side-effects as described above, unwarranted expectations due to media or company hype, the implant may fail and device support may end (e.g. due to company bankruptcy or newer products), image pre-processing to keep only a few selected edges may inadvertently filter out important safety cues, etc, etc.	Expectation management toward and by the (prospective) blind user. Expectation management is hampered by lack of population data on likely end results: training effort in relation to (perceived) benefit might prove unacceptable or inadequate. Also, blind users may misinterpret the visual content of soundscapes, so for their own safety they should never rely solely on The vOICe in mobile use but should keep relying on their cane, guide dog and/or trusted sighted guide. Lack of resources for 1-on-1 training support for best rehabilitation results is a concern.
Availability?	Unavailable at the time of writing. No FDA approval nor CE mark yet, and no patient has received a Blindsight implant as part of a clinical trial yet. Neuralink received Breakthrough Device designation from the FDA for the Blindsight brain implant, allowing for closer interaction with the FDA for feedback on medical device development. It is *not* FDA approval for market introduction. The US database ClinicalTrials.gov indicates that Neuralink is not yet recruiting blind people for a clinical trial (blindness is not among the listed medical conditions at the time of writing). Moreover, study inclusion conditions may imply that the implant is removed or turned off at the end of the trial. All in all it seems unlikely that a first Neuralink Blindsight implant trial will start before 2026, although Elon Musk hopes to do the first Neuralink Blindsight brain implant for a blind human later in 2025.	Globally available, and compatible with billions of existing devices. Meta should in due course lift their current policy constraints on using third-party apps with pixel-level camera access on their Meta Quest 2 and Quest 3 devices as well as their Meta Ray-Ban glasses, because these currently block blind people from using The vOICe sensory substitution on Meta devices (both Android app and web app). In the United States this may require lawsuits for blind accessibility based on the Americans with Disabilities Act (ADA) and Section 508. Privacy arguments do not apply because The vOICe works fine when blocking all camera view storage or broadcast at the system level. The vOICe merely locally transforms visual input into auditory output for blind accessibility of vision. It is unclear how restrictive the upcoming Meta Aria Gen 2 glasses will be, but the upcoming Google Android XR OS will reportedly allow full access to the camera view like on the smartphone.

Note that The vOICe and Neuralink Blindsight (and other brain implants) need not be mutually exclusive. A hybrid approach is conceivable where Neuralink Blindsight and The vOICe share the same camera view, such that the distorted and limited visual view of a Neuralink Blindsight brain implant can be supplemented with rich and undistorted visual feedback from The vOICe's soundscapes, including for instance the textures and shading missing from the edges-only Blindsight implant view. This combination may help the blind user to interpret the live camera view. Even OpenAI's ChatGPT agrees, but concludes that for the time being non-invasive sensory substitution will remain the superior choice for restoring functional vision to blind individuals. Furthermore, hyped super-human abilities such as "predator vision" by perceiving light outside the spectrum of visible light, e.g. infrared for thermal vision, are trivially dependent only on the choice of camera, and therefore apply equally to Neuralink Blindsight and The vOICe. Some outputs generated by other AI models such as Deepseek's Deepseek, Anthropic's Claude, xAI's Grok and Google's Gemini have been archived for future reference.

There are still many open questions regarding the role of neural plasticity: with blind adults, the brain's connectome ("physical wiring diagram") will be mostly fixed, but this need not apply to the brain-wide projectome (defining functional connectivity among brain areas): both brain implants and sensory substitution devices may lead to functional rewiring through changes in e.g.

dendritic spines, thereby modifying the projectome. Plenty of room for scientific studies with practical relevance far beyond blindness. With both brain implants and sensory substitution, visual feature detectors may be missing for properties such as curvature (corners), contour closure, overlap (occlusion) and containment (nesting), and to what extent these can be acquired through prolonged training is unknown. Vice versa, very few if any studies address possibilities for boosting top-down visual mental imagery, i.e., mental imagery engineering.

Other companies working on or toward brain implants for restoring vision include Cortigent (formerly Second Sight Medical Products, now a subsidiary of subsidiary, Orion brain implant), Sensor-Tech (ELVIS V brain implant), ReVision Implant (Occular brain implant), Phosphoenix, Science Corp, and INBRAIN, while still other brain implants for restoring vision are being developed in a more academic setting, such as the Intracortical Visual Prosthesis (ICVP) of Illinois Institute of Technology and the Gennaris bionic vision system of Monash University.

Investor note: Smart glasses with AI-based scene description, OCR for reading print, and visual-to-auditory sensory substitution for "raw vision" will erode the market opportunities for all implantable visual prostheses (retinal implants and brain implants for restoring vision to the blind). For Android devices, including smart glasses, AI-based scene description and OCR is already offered by for instance Google Lookout and Microsoft Seeing AI, while visual-to-auditory sensory substitution is offered by The vOICe for Android. Moreover, how does one convince the FDA that it is safe to use a high resolution visual cortical prosthesis for more than a short clinical trial, i.e., bring such a medical device to the market for daily (chronic) use? Run after numerous animal experiments a 5-year pilot study with a few human volunteers before potentially getting FDA approval with stringent post-market review requirements? Investors here need deep pockets and lots of patience for an uncertain and relatively small market. With The vOICe, by analogy with Linux, money may be made on training and support, and need not depend on poorly scalable proprietary hardware.

Some brain implant companies for their own convenience exclude sensory substitution devices from their BCI benchmarking publications and reporting to their investors, on the grounds that (by certain definitions) BCIs must interact directly with target neurons (e.g. in visual cortex), and not indirectly via other neurons (e.g. in auditory cortex). Similarly, some brain implant companies conveniently exclude sensory substitution devices from their reporting for not being medical devices. However, if it walks like an invasive BCI, talks like an invasive BCI, i.e. here offers much the same functional vision as an invasive BCI, it probably should be considered a BCI and visual prosthesis for all practical purposes - even if it non-invasively and indirectly activates or modulates target neurons in the brain. No hiding behind definitions and conventions.

Cf. An application-based taxonomy for brain–computer interfaces (Nature Biomedical Engineering, 2024), "We advocate that the term BCI should be used to define any technology that records brain activity and processes it on an electronic device, or any technology that stimulates brain activity based on computations performed on an electronic device, regardless of whether the technology is implanted or not." However, experts from different backgrounds are still struggling to come up with a good definition of what constitutes a BCI, as exemplified by BCI Society's 2024 BCI Working Definition that excludes the Dobelle and Orion brain implant for restoring vision just because these devices do not measure any brain activity. New proposal: "A BCI (Brain-Computer Interface) is a direct or indirect technological interface between the brain and a computing device, with the goal of substantially improving (restoring) communication and interactions between the brain and its physical environment and/or improving (restoring) functioning of the brain itself."

Finally, it should be stressed that BCIs in the form of cortical brain implants look very promising for restoring motor function, e.g. treating paralysis caused by ALS (amyotrophic lateral sclerosis) or spinal cord injury, or certain speech impairments, as well as for treating certain mental conditions such as depression, but their net benefit as a visual prosthesis remains questionable.

More about Neuralink Blindsight on the Neuralink website. May The vOICe be with you!

Note: bone conduction stereo earbuds can be hidden in the legs of smart glasses for the blind.

Copyright © 1996 - 2025 Peter B.L. Meijer

``Everything should be made as simple as possible, but not simpler''

quote attributed to Albert Einstein